

Calibration Register of Pulp Temperature Sensor

Container N°	MSGU 911304-0
---------------------	----------------------

Sensor N° 1			
Calibration	Reference Temperature	T° Sensor (°C)	Correction Factor Sensor N° 1
1	0.0	-0.2	+0.2
2	0.0	-0.2	
3	0.0	-0.2	
Sensor N° 2			
Calibration	Reference Temperature	T° Sensor (°C)	Correction Factor Sensor N° 2
1	0.0	-0.2	+0.2
2	0.0	-0.2	
3	0.0	-0.2	
Sensor N° 3			
Calibration	Reference Temperature	T° Sensor (°C)	Correction Factor Sensor N° 3
1	0.0	-0.2	+0.2
2	0.0	-0.2	
3	0.0	-0.2	

- The Correction Factor values were entered into the recorder and temperature Values in the treatment chart will be automatically adjusted.
- The Correction Factor values were NOT entered into the recorder and temperature Values in the treatment chart must be manually adjusted using the offset values.

CARLOS FERNANDEZ RIFFO,
INGENIERO AGRONOMO

**Name, Signature and Stamp
Authorized SAG Officer**

